

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

MAR 08 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>JAMES MICHELS</u>	ID# <u>6502</u> Date <u>2/28/10</u>
Agency <u>NORTH SLOPE BOROUGH POLICE</u>	Phone # <u>852-6111</u>
Instrument Location <u>BARROW-SQUAD ROOM</u>	
DataMaster cdm S/N <u>130129</u>	
Alco S/N <u>X301568</u>	Target Value <u>.084</u> High Pressure <u>900</u>
Alco Test Values	
<u>.084</u> 1 st Alco	<u>.084</u> 2 nd Alco
Signature <u>[Signature]</u> <u>SGT. MICHELS</u> <u>3/11/10</u>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of March, 2010.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland

Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130129

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

FEBRUARY 28, 2010

OPERATOR'S NAME:

MICHELSON JAMES/D

OPERATOR'S NUMBER: 6502

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: DMW

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .004

ALCO S/N: X301568

BREATH ANALYSIS

.004 ADJUSTED FOR 30.17 in		
ALCO TARGET	.004	04:04
BLANK TEST	.000	04:05
INTERNAL STANDARD	VERIFIED	04:05
ALCO TV 30.17 in	.004	04:06
BLANK TEST	.000	04:06
SUBJECT SAMPLE	.000	04:07
BLANK TEST	.000	04:08
ALCO TV 30.17 in	.004	04:08
BLANK TEST	.000	04:09

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

FEBRUARY 28, 2010

TIME 04:10

DIAGNOSTIC CHECK

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	40c
BREATH TUBE:	40c
BAROMETER:	30.17 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~"